
Midmark M11 - v1.0.0
Total Cycles: 2390

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2022 02:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.8 F 0.0
2:00 190.3 F 3.4
4:00 204.5 F 8.4
6:00 220.9 F 10.8
8:00 232.7 F 9.6
10:00 234.8 F 8.6
12:00 244.9 F 13.4
14:00 244.6 F 12.3
16:00 257.2 F 19.3
18:00 269.1 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.8 F 29.1
2:00 271.6 F 28.3
3:00 272.9 F 29.3
4:00 272.0 F 28.6

Min 271.5 F 28.3
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 233.7 F 0.0
10:00 229.8 F 0.0
15:00 233.5 F 0.0
20:00 223.1 F 0.0
25:00 226.8 F 0.0
30:00 231.1 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 19:08
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:58
CYCLE COMPLETE

10/03/2022 03:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____