
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001112
12:53 08/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:02
PRESSURIZING 4:02
STERILIZING 5:29
136.0 C 221kPa 5:29
136.4 C 224kPa 5:59
136.4 C 223kPa 6:29
136.2 C 223kPa 6:59
136.2 C 220kPa 7:29
136.3 C 224kPa 7:59
136.3 C 223kPa 8:29
136.2 C 223kPa 8:59
136.2 C 223kPa 8:59
VENTING 8:59
AIR DRYING 9:58
CYCLE COMPLETE 18:39

Digital Signature #
83C66FEABD104768
