
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007904
10:43 01/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:21
STERILIZING 3:23
136.5 C 224kPa 3:23
136.3 C 224kPa 3:53
136.1 C 222kPa 4:23
136.4 C 224kPa 4:53
136.4 C 224kPa 5:23
136.0 C 217kPa 5:53
136.4 C 223kPa 6:23
136.3 C 222kPa 6:53
136.3 C 222kPa 6:53
VENTING 6:53
AIR DRYING 7:43
CYCLE COMPLETE 27:02

Digital Signature #
9E2DC6CEBB755477
