
Midmark M11 - v1.0.0
Total Cycles: 3124

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2023 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.4 F 0.0
2:00 198.9 F 1.8
4:00 214.4 F 7.5
6:00 230.5 F 13.2
8:00 234.7 F 9.6
10:00 238.1 F 9.0
12:00 244.1 F 12.1
14:00 246.1 F 13.3
16:00 259.2 F 20.3
18:00 270.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.7
2:00 271.5 F 28.3
3:00 272.4 F 28.8
4:00 271.5 F 28.1

Min 271.3 F 28.1
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 217.9 F 0.0
10:00 233.9 F 0.0
15:00 238.3 F 0.0
20:00 232.6 F 0.0
25:00 222.9 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 2:24
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

04/18/2023 03:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____