
Midmark M11 - v1.0.0
Total Cycles: 4826

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2024 11:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.6 F 0.0
2:00 151.5 F 0.7
4:00 162.2 F 1.5
6:00 190.5 F 5.1
8:00 205.7 F 9.3
10:00 220.1 F 13.5
12:00 223.5 F 7.3
14:00 234.3 F 11.9
16:00 234.2 F 8.6
18:00 242.1 F 11.2
20:00 245.0 F 12.8
22:00 256.5 F 18.8
24:00 266.9 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.0 F 28.8
2:00 272.5 F 29.0
3:00 272.1 F 28.7
4:00 271.7 F 28.4

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.3 F 0.0
5:00 212.7 F 0.0
10:00 215.7 F 0.0
15:00 219.0 F 0.0
20:00 238.5 F 0.0
25:00 229.1 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 25:46
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:04:21
CYCLE COMPLETE

08/16/2024 12:41 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____