
Midmark M11 - v1.0.5
Total Cycles: 536

Name: RITTER2
Sterilizer ID: V2313228
11/10/2021 01:46 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.3 F 0.0
2:00 189.5 F 3.6
4:00 196.1 F 8.5
6:00 214.4 F 12.5
8:00 223.9 F 8.2
10:00 236.5 F 13.2
12:00 238.9 F 10.9
14:00 237.0 F 9.5
16:00 250.2 F 15.8
18:00 261.2 F 22.1
20:00 271.2 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 271.4 F 29.3
2:00 271.8 F 29.6
3:00 272.1 F 29.9
4:00 272.3 F 30.0

Min 270.9 F 28.9
Max 273.4 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.4 F 0.0
5:00 237.1 F 0.0
10:00 219.4 F 0.0
15:00 220.7 F 0.0
20:00 206.5 F 0.0
25:00 188.4 F 0.0
30:00 185.9 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

11/10/2021 02:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____