
Midmark M11 - v1.0.0
Total Cycles: 829

Name: RITTER 1
Sterilizer ID: V2178198
05/20/2021 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.4 F 0.0
2:00 151.2 F 1.4
4:00 163.8 F 4.2
6:00 184.9 F 7.9
8:00 204.1 F 12.3
10:00 225.1 F 6.0
12:00 233.3 F 11.2
14:00 233.2 F 8.1
16:00 242.8 F 11.6
18:00 245.0 F 12.5
20:00 256.8 F 18.9
22:00 267.6 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 271.6 F 28.8
2:00 272.8 F 29.4
3:00 273.2 F 29.4
4:00 271.6 F 28.7

Min 271.0 F 28.3
Max 273.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.2 F 0.0
5:00 231.1 F 0.0
10:00 188.6 F 0.0
15:00 192.4 F 0.0
20:00 193.5 F 0.0
25:00 196.3 F 0.0
30:00 198.0 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 23:40
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:06
CYCLE COMPLETE

05/20/2021 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____