
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000493
11:13 16/04/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:45
PRESSURIZING 2:56
STERILIZING 4:07
136.1 C 220kPa 4:07
136.4 C 233kPa 4:37
136.3 C 233kPa 5:07
136.2 C 231kPa 5:37
136.4 C 233kPa 6:07
136.4 C 232kPa 6:37
136.4 C 233kPa 7:07
136.2 C 231kPa 7:37
136.2 C 231kPa 7:37
VENTING 7:37
AIR DRYING 8:28
CYCLE COMPLETE 23:28

Digital Signature #
D815F0F00E3635E6
