
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001518
11:03 19/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:14
PRESSURIZING 2:19
STERILIZING 3:21
136.0 C 221kPa 3:21
136.4 C 225kPa 3:51
136.3 C 225kPa 4:21
136.4 C 226kPa 4:51
136.4 C 224kPa 5:21
136.4 C 226kPa 5:51
136.4 C 224kPa 6:21
136.4 C 225kPa 6:51
136.4 C 225kPa 6:51
VENTING 6:51
AIR DRYING 7:41
CYCLE COMPLETE 7:48

Digital Signature #
1FA474CA69C03116
