

-----  
Midmark M11 - v1.0.5  
Total Cycles: 370

Name: RITTER2  
Sterilizer ID: V2313228  
09/16/2021 08:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.2 F 0.0  
2:00 180.7 F 3.1  
4:00 186.1 F 7.5  
6:00 208.3 F 12.8  
8:00 222.3 F 7.7  
10:00 234.4 F 13.0  
12:00 238.6 F 10.8  
14:00 237.9 F 9.6  
16:00 251.1 F 16.2  
18:00 261.8 F 22.7  
20:00 272.6 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 272.3 F 29.4  
2:00 271.1 F 29.1  
3:00 272.6 F 29.9  
4:00 272.5 F 30.2

Min 271.0 F 28.8  
Max 273.2 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.0 F 0.0  
5:00 227.8 F 0.0  
10:00 227.8 F 0.0  
15:00 223.1 F 0.0  
20:00 208.5 F 0.0  
25:00 218.6 F 0.0  
30:00 216.0 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 20:24  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:59:13  
CYCLE COMPLETE

09/16/2021 09:52 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02