

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 003472  
8:55 19/04/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:46  
PRESSURIZING 3:47  
STERILIZING 4:57  
136.5 C 221kPa 4:57  
136.3 C 224kPa 5:27  
136.1 C 221kPa 5:57  
136.3 C 223kPa 6:27  
136.4 C 223kPa 6:57  
136.1 C 220kPa 7:27  
136.3 C 223kPa 7:57  
136.4 C 222kPa 8:27  
136.4 C 222kPa 8:27  
VENTING 8:27  
AIR DRYING 9:21  
CYCLE COMPLETE 9:23  
-----

Digital Signature #  
633D6F684A58AB0B  
-----