
Midmark M11 - v1.0.5
Total Cycles: 299

Name: RITTER2
Sterilizer ID: V2313228
08/23/2021 08:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.0 F 0.0
2:00 142.5 F 1.7
4:00 157.6 F 4.9
6:00 181.6 F 9.5
8:00 206.7 F 5.9
10:00 218.9 F 7.9
12:00 232.9 F 13.2
14:00 236.3 F 10.6
16:00 235.8 F 9.1
18:00 250.0 F 16.0
20:00 261.9 F 22.9
22:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 273.3 F 29.8
2:00 271.7 F 29.2
3:00 272.8 F 29.9
4:00 272.0 F 29.1

Min 271.6 F 29.1
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.9 F 0.0
5:00 205.1 F 0.0
10:00 214.0 F 0.0
15:00 208.6 F 0.0
20:00 201.7 F 0.0
25:00 204.1 F 0.0
30:00 209.4 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

08/23/2021 09:10 AM
ACCEPTED: X
REJECTED:
Operator: U02