
Midmark M11 - v1.0.5
Total Cycles: 2397

Name: RITTER2
Sterilizer ID: V2313228
05/23/2023 08:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.2 F 0.0
2:00 185.5 F 2.5
4:00 190.9 F 7.2
6:00 214.2 F 12.8
8:00 226.2 F 8.1
10:00 238.1 F 12.4
12:00 240.9 F 11.8
14:00 240.2 F 10.8
16:00 253.8 F 17.7
18:00 265.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.7
1:00 272.1 F 29.5
2:00 272.1 F 29.1
3:00 271.4 F 28.9
4:00 272.5 F 29.6

Min 271.3 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.7 F 0.0
5:00 221.5 F 0.0
10:00 237.0 F 0.0
15:00 241.5 F 0.0
20:00 232.8 F 0.0
25:00 220.6 F 0.0
30:00 216.6 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 19:31
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 00:58:46
CYCLE COMPLETE

05/23/2023 09:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____