

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4220

Name: RITTER2  
Sterilizer ID: V2313228  
12/17/2024 09:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.8 F 0.0  
2:00 70.7 F 0.0  
4:00 74.6 F 0.0  
6:00 103.4 F 0.1  
8:00 143.1 F 0.2  
10:00 168.2 F 2.6  
12:00 191.2 F 5.7  
14:00 206.6 F 9.6  
16:00 218.5 F 10.9  
18:00 226.4 F 8.3  
20:00 237.4 F 12.9  
22:00 237.4 F 10.3  
24:00 236.2 F 8.6  
26:00 248.4 F 14.9  
28:00 259.8 F 21.2  
30:00 270.2 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.8 F 29.6  
2:00 271.6 F 28.7  
3:00 272.7 F 29.7  
4:00 272.2 F 29.1

Min 271.5 F 28.7  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
5:00 198.3 F 0.0  
10:00 188.3 F 0.0  
15:00 187.4 F 0.0  
20:00 191.4 F 0.0  
25:00 194.5 F 0.0  
30:00 199.9 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 30:56  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:09:07  
CYCLE COMPLETE

12/17/2024 10:54 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_