
Midmark M11 - v1.0.5
Total Cycles: 327

Name: RITTER2
Sterilizer ID: V2313228
08/31/2021 02:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 126.6 F 0.0
2:00 128.3 F 1.4
4:00 146.9 F 4.0
6:00 172.6 F 7.7
8:00 194.6 F 11.6
10:00 217.8 F 5.3
12:00 230.5 F 9.8
14:00 234.6 F 9.9
16:00 238.8 F 11.5
18:00 237.7 F 9.8
20:00 249.9 F 15.9
22:00 260.9 F 22.1
24:00 270.9 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.4
1:00 272.7 F 29.6
2:00 272.2 F 30.1
3:00 271.7 F 29.4
4:00 272.4 F 30.1

Min 271.2 F 29.3
Max 273.0 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 222.8 F 0.0
10:00 196.5 F 0.0
15:00 189.3 F 0.0
20:00 173.5 F 0.0
25:00 179.7 F 0.0
30:00 177.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 25:27
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:03:60
CYCLE COMPLETE

08/31/2021 03:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____