
Midmark M11 - v1.0.0
Total Cycles: 4436

Name: RITTER 1
Sterilizer ID: V2178198
04/08/2024 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.9 F 0.0
2:00 105.6 F 0.9
4:00 115.5 F 3.0
6:00 140.8 F 5.9
8:00 168.4 F 9.4
10:00 197.3 F 12.1
12:00 208.1 F 5.3
14:00 219.1 F 9.3
16:00 231.3 F 12.2
18:00 234.7 F 9.7
20:00 237.6 F 9.6
22:00 245.1 F 13.1
24:00 256.5 F 19.1
26:00 266.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.3
2:00 271.7 F 28.6
3:00 271.5 F 28.9
4:00 272.5 F 29.2

Min 271.2 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.5 F 0.0
5:00 214.9 F 0.0
10:00 216.4 F 0.0
15:00 210.8 F 0.0
20:00 215.3 F 0.0
25:00 225.3 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 27:35
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:06:13
CYCLE COMPLETE

04/08/2024 04:21 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____