
Midmark M11 - v1.0.0
Total Cycles: 4571

Name: RITTER 1
Sterilizer ID: V2178198
05/22/2024 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 123.0 F 0.0
2:00 126.9 F 0.8
4:00 137.0 F 2.9
6:00 148.4 F 5.9
8:00 194.1 F 9.4
10:00 208.2 F 13.4
12:00 210.9 F 5.1
14:00 221.9 F 8.9
16:00 232.5 F 13.0
18:00 234.1 F 9.1
20:00 240.0 F 11.1
22:00 242.7 F 12.1
24:00 253.8 F 17.7
26:00 264.0 F 23.7
28:00 272.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.6
1:00 272.3 F 29.4
2:00 272.8 F 29.5
3:00 272.2 F 29.6
4:00 272.5 F 29.4

Min 271.0 F 28.5
Max 273.0 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 207.9 F 0.0
10:00 214.8 F 0.0
15:00 216.2 F 0.0
20:00 225.0 F 0.0
25:00 231.5 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 28:16
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:06:50
CYCLE COMPLETE

05/22/2024 10:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____