
Midmark M11 - v1.0.5
Total Cycles: 2484

Name: RITTER2
Sterilizer ID: V2313228
06/08/2023 09:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.9 F 0.0
2:00 105.5 F 0.0
4:00 123.3 F 0.0
6:00 154.3 F 2.5
8:00 176.6 F 5.9
10:00 196.0 F 9.5
12:00 216.9 F 11.4
14:00 222.7 F 8.0
16:00 234.1 F 12.8
18:00 236.0 F 10.1
20:00 236.1 F 8.8
22:00 248.0 F 14.9
24:00 259.6 F 21.4
26:00 270.4 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.6 F 29.8
2:00 271.3 F 29.1
3:00 272.6 F 29.9
4:00 271.8 F 29.3

Min 271.2 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.2 F 0.0
5:00 203.7 F 0.0
10:00 203.7 F 0.0
15:00 214.8 F 0.0
20:00 221.9 F 0.0
25:00 222.0 F 0.0
30:00 223.6 F 0.0

Drying Complete

FILLING: 0:49
HEATING: 26:51
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:04:54
CYCLE COMPLETE

06/08/2023 10:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____