

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2514

Name: RITTER 1  
Sterilizer ID: V2178198  
11/04/2022 08:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.1 F 0.0  
2:00 83.3 F 0.0  
4:00 100.2 F 0.2  
6:00 128.5 F 2.4  
8:00 153.9 F 5.8  
10:00 180.4 F 10.0  
12:00 199.9 F 6.3  
14:00 218.1 F 6.0  
16:00 229.3 F 10.0  
18:00 235.9 F 9.1  
20:00 240.3 F 11.0  
22:00 236.6 F 8.6  
24:00 251.2 F 15.5  
26:00 263.0 F 22.3  
28:00 271.6 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.0  
1:00 272.1 F 28.4  
2:00 272.7 F 28.6  
3:00 272.4 F 28.3  
4:00 272.0 F 28.0

Min 271.5 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 235.2 F 0.0  
Dry Cycle Aborted

FILLING: 0:49  
HEATING: 28:02  
STERILIZING: 4:00  
VENTING: 2:46  
DRYING: 2:21  
TOTAL CYCLE: 00:37:58

11/04/2022 09:23 AM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_