
Midmark M11 - v1.0.0
Total Cycles: 5449

Name: RITTER 1
Sterilizer ID: V2178198
03/26/2025 01:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.9 F 0.0
2:00 163.6 F 3.4
4:00 179.2 F 7.4
6:00 197.5 F 12.0
8:00 214.2 F 5.3
10:00 226.1 F 9.9
12:00 236.7 F 8.9
14:00 241.0 F 11.4
16:00 238.0 F 9.5
18:00 251.1 F 15.7
20:00 262.0 F 22.1
22:00 272.2 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.5 F 29.7
2:00 272.7 F 29.3
3:00 271.4 F 28.8
4:00 272.0 F 29.4

Min 271.3 F 28.8
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 210.8 F 0.0
10:00 230.5 F 0.0
15:00 238.8 F 0.0
20:00 235.7 F 0.0
25:00 210.9 F 0.0
30:00 207.7 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 23:24
STERILIZING: 4:00
VENTING: 3:54
DRYING: 30:00
TOTAL CYCLE: 01:02:28
CYCLE COMPLETE

03/26/2025 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____