
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000381
15:33 21/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:11
STERILIZING 3:09
136.0 C 228kPa 3:09
136.4 C 232kPa 3:39
136.2 C 232kPa 4:09
136.1 C 231kPa 4:39
136.4 C 233kPa 5:09
136.4 C 232kPa 5:39
136.1 C 231kPa 6:09
136.4 C 233kPa 6:39
136.4 C 233kPa 6:39
VENTING 6:39
AIR DRYING 7:30
CYCLE COMPLETE 7:37

Digital Signature #
E03F21123060E750
