
Midmark M11 - v1.0.5
Total Cycles: 2310

Name: RITTER2
Sterilizer ID: V2313228
04/28/2023 07:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.2 F 0.0
2:00 79.4 F 0.0
4:00 87.6 F 0.0
6:00 119.3 F 0.0
8:00 152.8 F 2.2
10:00 179.0 F 5.5
12:00 195.3 F 9.3
14:00 209.3 F 13.4
16:00 220.1 F 6.1
18:00 231.2 F 10.1
20:00 235.1 F 9.4
22:00 241.7 F 11.9
24:00 239.7 F 10.4
26:00 252.3 F 16.6
28:00 263.3 F 23.0
30:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.1 F 28.7
2:00 272.5 F 29.3
3:00 271.5 F 28.5
4:00 272.2 F 28.8

Min 271.4 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.1 F 0.0
5:00 192.1 F 0.0
10:00 207.8 F 0.0
15:00 204.9 F 0.0
20:00 200.1 F 0.0
25:00 197.7 F 0.0
30:00 196.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:12
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:08:10
CYCLE COMPLETE

04/28/2023 08:44 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____