
Midmark M11 - v1.0.5
Total Cycles: 537

Name: RITTER2
Sterilizer ID: V2313228
11/10/2021 02:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 183.9 F 3.7
4:00 199.8 F 9.1
6:00 218.0 F 8.2
8:00 230.8 F 10.8
10:00 233.2 F 8.7
12:00 239.3 F 10.4
14:00 247.1 F 14.4
16:00 259.2 F 21.1
18:00 269.9 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.5
2:00 271.8 F 29.7
3:00 271.9 F 29.2
4:00 272.2 F 29.9

Min 271.5 F 29.0
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 211.7 F 0.0
10:00 219.1 F 0.0
15:00 223.3 F 0.0
20:00 224.7 F 0.0
25:00 225.8 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 2:40
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

11/10/2021 03:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____