
Midmark M11 - v1.0.0
Total Cycles: 665

Name: MIDMARK
Sterilizer ID: V2178198
02/11/2021 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.7 F 0.0
2:00 132.7 F 0.6
4:00 141.5 F 2.5
6:00 184.0 F 6.4
8:00 201.3 F 11.9
10:00 213.1 F 5.5
12:00 227.1 F 10.7
14:00 229.7 F 7.0
16:00 244.4 F 13.5
18:00 243.3 F 11.8
20:00 256.3 F 18.5
22:00 267.9 F 25.6
24:00 271.9 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 271.8 F 28.6
2:00 271.7 F 28.5
3:00 271.9 F 28.6
4:00 272.4 F 28.8

Min 271.6 F 28.5
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 242.1 F 0.0
10:00 227.6 F 0.0
15:00 226.9 F 0.0
20:00 227.1 F 0.0
25:00 228.3 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 24:03
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 01:02:10
CYCLE COMPLETE

02/11/2021 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____