
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001163
8:40 11/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:29
STERILIZING 3:55
136.0 C 220kPa 3:55
136.4 C 224kPa 4:25
136.2 C 222kPa 4:55
136.4 C 225kPa 5:25
136.2 C 223kPa 5:55
136.4 C 223kPa 6:25
136.2 C 224kPa 6:55
136.2 C 223kPa 7:25
136.2 C 223kPa 7:25
VENTING 7:25
AIR DRYING 8:22
CYCLE COMPLETE 23:22

Digital Signature #
AB2450F573C83548
