

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001614  
13:30 20/11/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:04  
PRESSURIZING 3:04  
STERILIZING 4:08  
136.4 C 225kPa 4:08  
136.3 C 225kPa 4:38  
136.2 C 224kPa 5:08  
136.4 C 227kPa 5:38  
136.4 C 226kPa 6:08  
136.4 C 226kPa 6:38  
136.1 C 223kPa 7:08  
136.3 C 225kPa 7:38  
136.3 C 225kPa 7:38  
VENTING 7:38  
AIR DRYING 8:35  
CYCLE COMPLETE 8:39  
-----

Digital Signature #  
BE06990B495EE3BC  
-----