
Midmark M11 - v1.0.0
Total Cycles: 2983

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2023 04:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.9 F 0.0
2:00 182.3 F 3.3
4:00 189.4 F 7.9
6:00 232.0 F 12.2
8:00 225.8 F 8.6
10:00 240.5 F 11.1
12:00 242.8 F 11.8
14:00 239.4 F 10.4
16:00 253.8 F 17.2
18:00 265.3 F 23.9
20:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.3 F 28.8
2:00 272.6 F 29.2
3:00 271.9 F 29.1
4:00 272.2 F 28.8

Min 271.5 F 28.6
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 231.3 F 0.0
10:00 236.6 F 0.0
15:00 233.5 F 0.0
20:00 229.0 F 0.0
25:00 228.4 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

03/08/2023 05:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____