
Midmark M11 - v1.0.0
Total Cycles: 5117

Name: RITTER 1
Sterilizer ID: V2178198
11/20/2024 05:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.6 F 0.0
2:00 169.9 F 0.2
4:00 197.5 F 1.0
6:00 214.8 F 4.9
8:00 226.5 F 9.0
10:00 239.5 F 13.5
12:00 237.2 F 10.1
14:00 238.2 F 8.6
16:00 243.9 F 11.8
18:00 246.2 F 13.3
20:00 258.4 F 19.8
22:00 269.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.9 F 28.3
3:00 271.7 F 28.6
4:00 272.5 F 28.9

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.6 F 0.0
5:00 226.3 F 0.0
10:00 229.7 F 0.0
15:00 226.9 F 0.0
20:00 215.0 F 0.0
25:00 214.9 F 0.0
30:00 233.3 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:57
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:01:50
CYCLE COMPLETE

11/20/2024 06:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____