
Midmark M11 - v1.0.0
Total Cycles: 3721

Name: RITTER 1
Sterilizer ID: V2178198
09/05/2023 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.8 F 0.0
2:00 181.5 F 3.3
4:00 197.2 F 8.0
6:00 216.5 F 10.3
8:00 227.6 F 8.4
10:00 238.0 F 11.6
12:00 241.3 F 11.5
14:00 239.8 F 10.2
16:00 253.8 F 17.1
18:00 265.6 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.4 F 28.6
2:00 271.8 F 28.4
3:00 272.3 F 28.5
4:00 271.7 F 28.4

Min 271.5 F 28.0
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 210.9 F 0.0
10:00 224.1 F 0.0
15:00 229.4 F 0.0
20:00 224.9 F 0.0
25:00 221.2 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

09/05/2023 05:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____