
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002473
10:27 29/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:52
STERILIZING 3:58
136.0 C 220kPa 3:58
136.2 C 223kPa 4:28
136.2 C 223kPa 4:58
136.3 C 224kPa 5:28
136.4 C 222kPa 5:58
136.4 C 223kPa 6:28
136.1 C 223kPa 6:58
136.3 C 225kPa 7:28
136.3 C 225kPa 7:28
VENTING 7:28
AIR DRYING 8:21
CYCLE COMPLETE 38:21

Digital Signature #
C3B314947A2610DC
