
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003371
13:37 01/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:27
PRESSURIZING 2:32
STERILIZING 3:28
136.4 C 224kPa 3:28
136.5 C 225kPa 3:58
136.2 C 223kPa 4:28
136.4 C 223kPa 4:58
136.4 C 224kPa 5:28
136.1 C 222kPa 5:58
136.1 C 222kPa 6:28
136.3 C 222kPa 6:58
136.3 C 222kPa 6:58
VENTING 6:58
AIR DRYING 7:53
CYCLE COMPLETE 7:54

Digital Signature #
54A60A1B7091B1B7
