
Midmark M11 - v1.0.0
Total Cycles: 3662

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2023 10:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.2 F 0.0
2:00 168.8 F 2.5
4:00 173.4 F 5.7
6:00 194.0 F 9.7
8:00 215.3 F 7.8
10:00 223.7 F 7.2
12:00 231.3 F 11.7
14:00 233.7 F 8.1
16:00 245.0 F 12.8
18:00 243.6 F 12.0
20:00 254.9 F 17.9
22:00 265.2 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.7
1:00 272.4 F 29.1
2:00 272.0 F 28.8
3:00 271.5 F 28.6
4:00 271.5 F 28.6

Min 271.4 F 28.6
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 232.0 F 0.0
10:00 218.4 F 0.0
15:00 199.1 F 0.0
20:00 194.3 F 0.0
25:00 202.9 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

08/16/2023 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____