
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000300
13:44 05/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:59
STERILIZING 3:58
136.0 C 220kPa 3:58
136.3 C 220kPa 4:28
136.1 C 223kPa 4:58
136.4 C 225kPa 5:28
136.4 C 226kPa 5:58
136.4 C 225kPa 6:28
136.4 C 225kPa 6:58
136.4 C 225kPa 7:28
136.4 C 225kPa 7:28
VENTING 7:28
AIR DRYING 8:18
CYCLE COMPLETE 8:59

Digital Signature #
8A00C1BB0BFC94D2
