
Midmark M11 - v1.0.0
Total Cycles: 1824

Name: RITTER 1
Sterilizer ID: V2178198
04/14/2022 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.2 F 0.0
2:00 105.4 F 0.0
4:00 130.5 F 0.0
6:00 157.8 F 1.8
8:00 181.3 F 5.6
10:00 196.3 F 9.4
12:00 212.9 F 11.9
14:00 222.2 F 7.0
16:00 235.5 F 11.4
18:00 233.8 F 7.9
20:00 242.8 F 12.2
22:00 243.8 F 12.3
24:00 256.3 F 18.9
26:00 267.5 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.1
2:00 271.7 F 28.7
3:00 272.7 F 29.2
4:00 271.7 F 28.5

Min 271.5 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 205.6 F 0.0
10:00 222.3 F 0.0
15:00 225.1 F 0.0
20:00 223.6 F 0.0
25:00 225.6 F 0.0
30:00 227.4 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 27:21
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:05:21
CYCLE COMPLETE

04/14/2022 11:19 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____