

-----  
Midmark M11 - v1.0.5  
Total Cycles: 159

Name: RITTER2  
Sterilizer ID: V2313228  
07/05/2021 10:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 207.5 F 0.0  
2:00 202.4 F 3.8  
4:00 208.6 F 9.7  
6:00 221.9 F 7.3  
8:00 233.7 F 11.4  
10:00 236.3 F 9.8  
12:00 237.4 F 9.3  
14:00 250.9 F 16.2  
16:00 262.9 F 23.2  
18:00 272.3 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.2 F 29.3  
2:00 272.2 F 29.7  
3:00 271.7 F 28.9  
4:00 272.9 F 29.6  
  
Min 271.4 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
5:00 233.0 F 0.0  
10:00 222.0 F 0.0  
15:00 218.3 F 0.0  
20:00 213.4 F 0.0  
25:00 211.9 F 0.0  
30:00 209.7 F 0.0  
Drying Complete

FILLING: 2:18  
HEATING: 18:10  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:38  
CYCLE COMPLETE

07/05/2021 11:26 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01