
Midmark M11 - v1.0.5
Total Cycles: 1616

Name: RITTER2
Sterilizer ID: V2313228
10/18/2022 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.6 F 0.0
2:00 198.6 F 3.6
4:00 200.1 F 8.9
6:00 221.9 F 8.1
8:00 234.6 F 10.8
10:00 234.1 F 8.7
12:00 239.8 F 10.3
14:00 247.9 F 14.7
16:00 260.2 F 21.6
18:00 271.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.9
2:00 272.1 F 29.3
3:00 271.3 F 28.9
4:00 272.2 F 29.7

Min 271.3 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 228.4 F 0.0
10:00 223.7 F 0.0
15:00 230.0 F 0.0
20:00 231.4 F 0.0
25:00 229.1 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 3:25
HEATING: 18:41
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

10/18/2022 12:52 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____