

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4568

Name: RITTER 1  
Sterilizer ID: V2178198  
05/21/2024 11:01 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 144.5 F 0.0  
2:00 146.3 F 2.0  
4:00 154.8 F 5.2  
6:00 178.8 F 8.9  
8:00 199.3 F 13.1  
10:00 214.0 F 6.7  
12:00 225.3 F 11.0  
14:00 228.2 F 6.7  
16:00 238.7 F 11.9  
18:00 237.3 F 9.6  
20:00 249.3 F 15.5  
22:00 259.8 F 21.5  
24:00 269.7 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.1  
1:00 272.7 F 29.8  
2:00 271.9 F 29.2  
3:00 271.4 F 29.3  
4:00 272.6 F 29.8

Min 271.1 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.8 F 0.0  
5:00 215.0 F 0.0  
10:00 224.7 F 0.0  
15:00 229.2 F 0.0  
20:00 227.6 F 0.0  
25:00 230.4 F 0.0  
30:00 230.6 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 25:03  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:03:37  
CYCLE COMPLETE

05/21/2024 12:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_