
Midmark M11 - v1.0.0
Total Cycles: 4116

Name: RITTER 1
Sterilizer ID: V2178198
12/14/2023 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.2 F 0.0
2:00 162.5 F 1.5
4:00 169.2 F 4.4
6:00 191.2 F 8.7
8:00 211.2 F 13.4
10:00 224.0 F 6.3
12:00 234.5 F 10.7
14:00 233.0 F 7.4
16:00 244.7 F 12.7
18:00 242.0 F 10.9
20:00 254.0 F 17.0
22:00 264.7 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.6 F 28.7
2:00 272.6 F 28.8
3:00 271.9 F 28.5
4:00 271.9 F 28.2

Min 270.9 F 27.8
Max 272.9 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 203.9 F 0.0
10:00 200.0 F 0.0
15:00 197.5 F 0.0
20:00 192.7 F 0.0
25:00 191.1 F 0.0
30:00 191.1 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 23:54
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:02:58
CYCLE COMPLETE

12/14/2023 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____