

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001899  
11:00 26/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:59  
STERILIZING 3:02  
136.0 C 222kPa 3:02  
136.0 C 222kPa 3:32  
136.1 C 223kPa 4:02  
136.4 C 226kPa 4:32  
136.4 C 225kPa 5:02  
136.4 C 225kPa 5:32  
136.4 C 226kPa 6:02  
136.4 C 226kPa 6:32  
136.4 C 226kPa 6:32  
VENTING 6:32  
AIR DRYING 7:24  
CYCLE COMPLETE 13:58

-----  
Digital Signature #  
154F63BAC996D94F  
-----