
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000974
13:09 09/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:45
STERILIZING 3:54
136.4 C 233kPa 3:54
136.4 C 233kPa 4:24
136.4 C 233kPa 4:54
136.3 C 232kPa 5:24
136.2 C 232kPa 5:54
136.4 C 233kPa 6:24
136.4 C 233kPa 6:54
136.4 C 233kPa 7:24
136.4 C 233kPa 7:24
VENTING 7:24
AIR DRYING 8:16
CYCLE COMPLETE 8:19

Digital Signature #
6FDD23B747D0508D
