
Midmark M11 - v1.0.0
Total Cycles: 3063

Name: RITTER 1
Sterilizer ID: V2178198
03/31/2023 09:02 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.5 F 0.0
2:00 176.3 F 1.3
4:00 186.1 F 6.5
6:00 209.8 F 12.5
8:00 225.4 F 7.7
10:00 237.0 F 10.1
12:00 244.2 F 12.7
14:00 243.3 F 11.7
16:00 257.8 F 19.3
18:00 270.3 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 27.9
1:00 272.1 F 28.0
2:00 272.8 F 28.6
3:00 272.2 F 28.5

Min 271.4 F 27.7
Max 272.9 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 226.9 F 0.0
Dry Cycle Aborted

FILLING: 1:12
HEATING: 18:45
STERILIZING: 3:00
VENTING: 2:50
DRYING: 3:59
TOTAL CYCLE: 00:29:47

03/31/2023 09:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____