

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1710

Name: RITTER 1  
Sterilizer ID: V2178198  
03/14/2022 08:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.3 F 0.0  
2:00 165.0 F 1.4  
4:00 185.6 F 5.1  
6:00 199.0 F 9.9  
8:00 212.7 F 5.1  
10:00 226.2 F 8.6  
12:00 236.4 F 11.9  
14:00 238.9 F 10.8  
16:00 236.5 F 8.8  
18:00 251.0 F 15.7  
20:00 262.9 F 22.5  
22:00 272.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.4 F 29.1  
2:00 271.7 F 28.4  
3:00 271.8 F 28.9  
4:00 272.4 F 28.9

Min 271.3 F 28.2  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.9 F 0.0  
5:00 220.6 F 0.0  
10:00 238.7 F 0.0  
15:00 236.6 F 0.0  
20:00 231.3 F 0.0  
25:00 222.0 F 0.0  
30:00 220.6 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 22:16  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:00:15  
CYCLE COMPLETE

03/14/2022 09:27 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_