
Midmark M11 - v1.0.5
Total Cycles: 3881

Name: RITTER2
Sterilizer ID: V2313228
08/21/2024 03:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.0 F 0.0
2:00 183.1 F 3.8
4:00 210.9 F 9.2
6:00 222.2 F 6.8
8:00 233.6 F 11.2
10:00 234.9 F 9.1
12:00 238.8 F 9.7
14:00 249.2 F 15.2
16:00 261.5 F 22.3
18:00 272.7 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.7 F 29.6
2:00 271.7 F 28.8
3:00 272.2 F 29.7
4:00 272.3 F 29.3

Min 271.2 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.4 F 0.0
5:00 214.6 F 0.0
10:00 205.6 F 0.0
15:00 202.5 F 0.0
20:00 200.9 F 0.0
25:00 200.1 F 0.0
30:00 203.1 F 0.0

Drying Complete

FILLING: 1:60
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

08/21/2024 04:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____