
Midmark M11 - v1.0.5
Total Cycles: 2584

Name: RITTER2
Sterilizer ID: V2313228
06/29/2023 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 76.0 F 0.0
4:00 83.9 F 0.0
6:00 99.0 F 0.0
8:00 131.7 F 1.8
10:00 161.0 F 4.5
12:00 185.4 F 7.6
14:00 201.5 F 11.1
16:00 213.8 F 4.5
18:00 224.8 F 8.5
20:00 235.8 F 13.2
22:00 237.2 F 10.6
24:00 235.1 F 8.8
26:00 249.6 F 15.4
28:00 260.4 F 21.8
30:00 270.8 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.5 F 29.7
2:00 271.5 F 29.1
3:00 272.6 F 29.7
4:00 271.6 F 29.0

Min 271.4 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.8 F 0.0
5:00 198.6 F 0.0
10:00 203.5 F 0.0
15:00 204.8 F 0.0
20:00 201.0 F 0.0
25:00 201.1 F 0.0
30:00 204.4 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:45
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:08:45
CYCLE COMPLETE

06/29/2023 08:24 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____