

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1829

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2022 09:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 165.1 F 0.0  
2:00 168.0 F 0.0  
4:00 174.2 F 3.7  
6:00 196.3 F 9.1  
8:00 221.6 F 7.7  
10:00 231.4 F 9.9  
12:00 235.5 F 8.6  
14:00 244.4 F 12.9  
16:00 244.4 F 13.0  
18:00 257.6 F 20.0  
20:00 269.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.8 F 29.6  
2:00 271.6 F 28.9  
3:00 272.7 F 29.8  
4:00 272.2 F 29.2

Min 271.5 F 28.8  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
5:00 205.8 F 0.0  
10:00 223.2 F 0.0  
15:00 229.1 F 0.0  
20:00 224.2 F 0.0  
25:00 225.0 F 0.0  
30:00 223.9 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 21:02  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:09  
CYCLE COMPLETE

12/13/2022 10:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_