
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001853
13:11 23/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:46
STERILIZING 4:47
136.0 C 220kPa 4:47
136.3 C 223kPa 5:17
136.2 C 222kPa 5:47
136.4 C 224kPa 6:17
136.4 C 223kPa 6:47
136.2 C 223kPa 7:17
136.4 C 224kPa 7:47
136.4 C 225kPa 8:17
136.4 C 225kPa 8:17
VENTING 8:17
AIR DRYING 9:11
CYCLE COMPLETE 9:13

Digital Signature #
9C700AD6D18988C9
