

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1436

Name: RITTER 1  
Sterilizer ID: V2178198  
12/08/2021 04:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 163.9 F 0.0  
2:00 170.4 F 2.8  
4:00 189.3 F 6.8  
6:00 213.6 F 12.2  
8:00 223.2 F 6.5  
10:00 236.4 F 11.8  
12:00 234.7 F 8.7  
14:00 241.1 F 10.1  
16:00 246.6 F 13.5  
18:00 259.1 F 20.2  
20:00 270.4 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.5 F 28.8  
2:00 272.1 F 28.9  
3:00 272.1 F 28.6  
4:00 272.1 F 29.0  
  
Min 271.5 F 28.3  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.4 F 0.0  
5:00 238.3 F 0.0  
10:00 228.9 F 0.0  
15:00 226.5 F 0.0  
20:00 228.7 F 0.0  
25:00 229.2 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:52  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:16  
CYCLE COMPLETE

12/08/2021 05:02 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_