
Midmark M11 - v1.0.0
Total Cycles: 1408

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2021 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 174.8 F 3.3
4:00 191.6 F 7.3
6:00 207.3 F 12.4
8:00 227.6 F 6.7
10:00 238.7 F 12.0
12:00 236.2 F 9.1
14:00 238.6 F 9.5
16:00 248.3 F 14.3
18:00 261.1 F 21.4
20:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.2
1:00 272.7 F 29.2
2:00 271.9 F 28.4
3:00 272.7 F 29.2
4:00 271.8 F 28.4

Min 271.0 F 28.0
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 233.8 F 0.0
10:00 206.3 F 0.0
15:00 206.1 F 0.0
20:00 204.2 F 0.0
25:00 212.4 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

11/30/2021 02:09 PM
ACCEPTED: X
REJECTED:
Operator: U02