

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3181

Name: RITTER2  
Sterilizer ID: V2313228  
12/04/2023 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 102.0 F 0.0  
2:00 105.2 F 0.3  
4:00 127.1 F 2.0  
6:00 158.8 F 5.2  
8:00 185.8 F 8.9  
10:00 205.9 F 13.3  
12:00 215.8 F 6.5  
14:00 226.4 F 10.9  
16:00 229.4 F 7.3  
18:00 241.3 F 12.9  
20:00 240.8 F 11.6  
22:00 252.8 F 17.6  
24:00 263.9 F 24.2  
26:00 271.9 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.8 F 30.1  
2:00 271.8 F 29.4  
3:00 271.6 F 29.7  
4:00 273.0 F 30.0

Min 271.1 F 29.2  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.2 F 0.0  
5:00 202.9 F 0.0  
10:00 195.0 F 0.0  
15:00 189.2 F 0.0  
20:00 185.9 F 0.0  
25:00 183.3 F 0.0  
30:00 181.7 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 26:10  
STERILIZING: 4:00  
VENTING: 3:35  
DRYING: 30:00  
TOTAL CYCLE: 01:04:40  
CYCLE COMPLETE

12/04/2023 03:47 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_