
Midmark M11 - v1.0.5
Total Cycles: 2781

Name: RITTER2
Sterilizer ID: V2313228
08/17/2023 08:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.6 F 0.0
2:00 161.2 F 0.0
4:00 181.6 F 0.0
6:00 206.0 F 0.5
8:00 221.1 F 5.0
10:00 232.8 F 9.4
12:00 239.4 F 11.2
14:00 242.1 F 12.0
16:00 240.7 F 10.9
18:00 239.3 F 10.3
20:00 253.4 F 17.3
22:00 265.1 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.4 F 28.8
2:00 272.8 F 29.4
3:00 271.7 F 28.6
4:00 272.6 F 29.0

Min 271.7 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 196.2 F 0.0
10:00 191.1 F 0.0
15:00 190.0 F 0.0
20:00 188.9 F 0.0
25:00 196.7 F 0.0
30:00 203.4 F 0.0
Drying Complete

FILLING: 0:39
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:01:38
CYCLE COMPLETE

08/17/2023 09:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____